



Welcome to the next step in your pet's health journey. By using the Holter monitor, you're helping to capture critical insights into your pet's heart function in the comfort of their home. This easy-to-use device ensures we can monitor their well-being without disrupting their daily routine, giving you peace of mind and them the care they deserve.

ATRIO
OCTOPUS
A SMALL DEVICE FOR BIG HEARTS

POWERED BY
ALERTE
DIGITAL VET

LOVE THEM. PROTECT THEM. **ALWAYS.**

Because your furry family deserves the best.



Instructions for Use

Thank you for helping us monitor your pet's heart health. Please follow these instructions carefully to ensure accurate results and your pet's comfort during the monitoring period.

DEVICE SETUP AND PLACEMENT

Your pet has four leads connected to electrodes that were securely attached by your vet after shaving small areas of fur. The leads are connected to the Atrio Octopus Holter device, which has been wrapped with self-sticking bandage tape and covered with a lightweight jacket or rashie to keep everything in place. Please do not attempt to adjust or remove the setup.

WATER PRECAUTIONS

The Holter device is **not waterproof**. Ensure your pet avoids baths, swimming, heavy rain, or any other situation where the device could get wet.

RECORDING PERIOD

- The device has already been turned on (indicated by a pulsing green light), so no additional action is required on your part.
- The device has been pre-programmed by your vet to record heart activity continuously for 24 hours and will stop pulsing and turn off on its own.

ACTIVITY GUIDELINES

- Keep your pet's activity level as low as possible. High-energy activities like running, jumping, or rough play should be avoided to prevent the leads or electrodes from disconnecting.
- Gentle, short walks are fine if necessary, but prioritise rest and low-movement activities.
- Monitor your pet for signs of discomfort or irritation caused by the device, bandages, or jacket. Contact your vet if you notice anything unusual.

HOLTER DIARY

Record significant events during the monitoring period, including:

- Mealtimes
- Bathroom breaks
- Rest or sleep periods
- High-energy activities
- Any unusual behaviour (e.g., scratching at the device or sudden lethargy).

This diary helps us align these events with the ECG readings for more accurate analysis.

CHECK THE SETUP REGULARLY

Periodically ensure the bandages, jacket, and device remain secure. If anything becomes loose or out of place, contact your vet or our technical support team.

NIGHT-TIME MONITORING

Create a quiet, low-activity environment during sleep. Ensure your pet is resting in a comfortable position that does not pull on the device or leads.

URGENT SITUATIONS

If the device unexpectedly stops working, becomes detached, or shows any signs of malfunction, do not attempt to fix it yourself. Contact your vet or our technical support team immediately for assistance.

POST-MONITORING

After the 24-hour recording period is complete, return the device to your vet as instructed. The vet will safely remove the device, leads, electrodes and bandages.

ATRIO
OCTOPUS
A SMALL DEVICE FOR BIG HEARTS

POWERED BY
ALERTE
DIGITAL VET



SUPPORT

For any concerns or issues during the monitoring period, please contact our technical support team at **0418 250 292**.



alertedv.com



info@alertedv.com

